



**ENVIRONMENTAL PROTECTION DIVISION**

**Richard E. Dunn, Director**

---

**Air Protection Branch**  
4244 International Parkway  
Suite 120  
Atlanta, Georgia 30354  
404-363-7000

June 4, 2019

**Amendment Number: 2**

Jerry Jeffers, Imaging Manager  
CHI Memorial Georgia  
100 Gross Crescent Circle  
Fort Oglethorpe, Georgia 30742

Subject: **Compliance Letter**  
License GA 1873-1

This letter refers to the inspection that was conducted on May 22, 2019, by me. At the conclusion of the inspection of activities authorized by Georgia Radioactive Material License **GA 1873-1**, a discussion of the findings was held with you.

The inspection was an examination of the activities conducted under the license as they relate to radiation safety and to compliance with the Georgia Department of Natural Resources "Rules and Regulations for Radioactive Materials," Chapter 391-3-17, and the conditions of the license. The inspection consisted of selective examinations of procedures and representative records, independent measurements and observations by the Department. One noted deficiency: Linearity, Accuracy, and Geometry had not been performed on the dose calibrator prior to use. (391-3-17-.05(29)(c)). A deadline of documentation was established.

Based on the documentation provided dated May 28, 2019, Geometry has been performed, linearity was also performed at the beginning of May. Therefore, no further action is warranted at this time.

If you have any questions, please do not hesitate to contact our office at (404) 362-2675.

Sincerely,

A handwritten signature in cursive script that reads "S. Walker".

Shatavia Walker  
Environmental Compliance Specialist  
Radioactive Materials Program